

CERTIFICATE OF ANALYSIS


Product Information

Product Name:	ToxinSensor™ Chromogenic LAL Endotoxin Assay Kit
Cat. No.:	L00350
Lot No.:	C50101804
Exp.Date:	01 Mar 2020

Kit Contents

Kit Components	Lot.No	Storage Condition
LAL Reagent Water	C50091804	Store at 2-8°C
Limulus Amebocyte Lysate (LAL)	C50011804	Store at 2-8°C
<i>E. coli</i> Endotoxin Standard	C50021804	Store at 2-8°C
Chromogenic Substrate	C50031804	Store at 2-8°C
Buffer S for Color-stabilizer #1	C50081804	Store at 2-8°C
Color-stabilizer #1	C50041804	Store at 2-8°C
Color-stabilizer #2	C50051804	Store at 2-8°C
Color-stabilizer #3	C50061804	Store at 2-8°C
Endotoxin-free tubes	C50291803	Store at 2-8°C
Endotoxin-free Tips, 200 µl	---	Store at 2-8°C
Endotoxin-free Tips, 1000 µl	C50301803	Store at 2-8°C
Incubation Rack	---	Store at 2-8°C

Endpoint Chromogenic LAL Assay



Test		Results
Sensitivity (OD 545nm)	Blank	0.089
	Maxi Endotoxin Standards	1.317
Linearity	Correlation Coefficient (R ²)	0.9959
	Endotoxin-free Tips, 1000 µl	0.005
Endotoxin-free materials (EU/ml)	Endotoxin-free tubes	0.004

Caution:

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by:  Date: 05/01/2018
Product Department