

CERTIFICATE OF ANALYSIS


Product Information

Product Name:	ToxinSensor™ Chromogenic LAL Endotoxin Assay Kit
Cat. No.:	L00350
Lot No.:	C51002004
Exp.Date:	01 Apr 2022
Storage Condition:	Store at 2-8°C
Quality Release Date:	04 Nov 2020

Kit Contents


Kit Components	Lot.No
LAL Reagent Water	C50072004
Limulus Amebocyte Lysate (LAL)	C50012004
<i>E. coli</i> Endotoxin Standard	C50022004
Chromogenic Substrate	C50032004
Color-stabilizer #1	C50042004
Color-stabilizer #2	C50052004
Color-stabilizer #3	C50062004
Endotoxin-free tubes	C50291904/C50292010
Endotoxin-free Tips, 200 µL	---
Endotoxin-free Tips, 1000 µL	C50302001
Incubation Rack	---

Standard Curve (0.01-0.1 EU/mL)



Test Items	Specifications	Results
0.01-0.1EU/mL(OD545 nm)	0 EU/mL(OD545 nm)≤0.30	0.0444
	0.1 EU/mL(OD545 nm) ≥0.50	0.5698
	0.01 EU/mL(OD545 nm)>0 EU/mL(OD545 nm)	Pass
	R ≥0.98; R ² ≥0.96	Pass

Standard Curve (0.1-1 EU/mL)



Test Items	Specifications	Results
0.1-1 EU/mL(OD545 nm)	0 EU/mL(OD545 nm) ≤0.2	0.0436
	1 EU/mL(OD545 nm) ≥0.75	0.7709
	0.1 EU/mL(OD545 nm)>0 EU/mL(OD545 nm)	Pass
	R ≥0.98; R ² ≥0.96	Pass
Endotoxin-free Tips, 1000 μL	<0.005 EU/mL	Pass
Endotoxin-free tubes	<0.005 EU/mL	Pass

Xiyang Ma

Quality Department

Caution:

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.