

Sample Name:MOG(35-55)
 Cat No. :RP10245
 Lot No. :E5154621A
 Time Processed :2:19:23
 Month-Day-Year Processed :01/07/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	4.787	195580	28592	0.761
2	11.013	10935	1924	0.043
3	12.107	2287	279	0.009
4	12.347	21636	2170	0.084
5	12.884	285265	22151	1.110
6	13.085	231071	24529	0.900
7	13.342	24484622	2466172	95.315
8	13.596	269035	38856	1.047
9	13.781	187645	17396	0.730
Total		25688075	2602069	100.000