


Mass Spectrum



Sample Information
Month-Day Processed : 08/11/20
Time Processed : 20:35:09
Injection Volume : 0.3
Sample Name : Flagelin22
Cat No. : RP19986
Lot No. : E6829620H
Theoretical MW : 2272.50
Observed MW : 2271.9

Interface : ESI
Nebulizing Gas Flow : 1.5L/min
CDL Temp : 250
Block Temp : 200

Equipment : GK11010007
Interface Bias : +4.5kV
Drying Gas Flow : 5 L/min
T.Flow : 0.2ml/min
B.conc : 50%H2O/50%MeOH