


Mass Spectrum



Sample Information
Month-Day Processed : 03/14/21
Time Processed : 12:49:16 AM
Injection Volume : 0.5
Sample Name : Flagelin22
Cat No. : RP19986
Lot No. : E8076621C
Theoretical MW : 2272.50
Observed MW : 2271.9

Interface : ESI
Nebulizing Gas Flow : 1.5L/min
CDL Temp : 250
Block Temp : 200
Equipment : GK11010007
Interface Bias : +4.5 kV
Drying Gas Flow : 5 L/min
T.Flow : 0.2 ml/min
B.conc : 50% H2O/50% MeOH